

## Numerics

### 6890 GC

- Communication with 3395A/3396B
  - Integrator, 1010-6, 1010-11, 1015-5, 1015-9
- Communication with 3395B/3396B
  - Integrator, 1010-7, 1015-6
- Communication with 3396B/C INET
  - Integrator, 1010-5, 1010-10
- Communication with Automatic Liquid Sampler, 1010-2, 1010-3, 1010-4, 1010-5, 1010-6, 1010-7, 1010-8, 1015-2, 1015-3, 1015-4, 1015-5, 1015-6, 1015-7, 1015-14
- Communication with A/D converter, 1010-8, 1015-7
- Communication with ChemStation, 1010-2, 1010-3, 1010-9, 1015-2, 1015-3, 1015-8, 1015-14
- Communication with Headspace sampler 7694, 1010-9, 1010-10, 1010-11, 1010-12, 1015-8, 1015-9, 1015-10
- Communication with Mass Selective Detector, 1010-3, 1015-3, 1015-14
- Communication with Modem, 1010-5
- Communication with Non-Agilent data system, 1010-4, 1015-4
- Communication with other instruments, 1010-13, 1015-11
- G1900A Purge and Tap, 1010-14, 1015-13
- GPIO cables, 8120-3446, 1010-2, 1010-3, 1010-9
- Grounding, 530-1

## A

### AC power board

- Circuitry diagram, 1230-7
- Connector circuitry diagram, 1220-6
- Connectors, 1230-5

### AC Power Supply

- AC board IPB, 1360-4
- IPB, 1360-2

### Activated charcoal trap, 520-8

### Active element

- NPD, Replacement, 320-13

### Actuator

- Air Requirements, 520-13

### Actuators

### Alignment, 1130-10

### Driver installation, 1130-5

### Drivers, 1130-4

### Installation, 1130-1

### IPB, 1350-6

### Address Error, 860-1

### ALS interface board, 430-23

### Analog

#### Cable, 1020-1

#### Connector, 1210-2

### Analog cables

#### 03396-60530,, 1010-10

#### G1530-60560, 1010-12, 1015-10

#### G1530-60570, 1010-6, 1010-7, 1010-8, 1010-11, 1015-5, 1015-6, 1015-7, 1015-9

### Analog out data loss, 830-3

### Analog out #1, 850-4

### Analog out #2, 850-4

### Analog-to-Digital Converter, 35900, 510-5

#### Communication with 6890 GC, 1010-8, 1015-7

#### Communication with ALS, 1010-8, 1015-7

### APG remote cables

#### 03396-61010, 1010-7, 1010-11, 1015-6, 1015-9

#### 03396-61020, 1010-6, 1010-11, 1015-5, 1015-9

#### 35900-60670, 1010-4, 1010-12, 1015-4, 1015-10

#### G1290-60570, 1010-9, 1010-10, 1010-11, 1010-12, 1015-8, 1015-9, 1015-10

#### G1530-60930, 1010-2, 1010-3, 1010-4, 1010-5, 1010-6, 1010-7, 1010-8, 1015-3, 1015-7, 1015-14

### APG remote connector, 1210-4, 1215-7

### APG remote start-stop cable

#### G1530-60930, 1010-2, 1010-3, 1010-4, 1010-5, 1010-6, 1010-7, 1015-3

### Auto flow zero, 130-1

### Automatic Liquid Sampler

#### Cable, 1020-3

#### Communication with 3395A/3396B

##### Integrator, 1010-6, 1015-5

#### Communication with 3395B/3396C

##### Integrator, 1010-7, 1015-6

#### Communication with 6890 GC, 1010-2,

##### 1010-3, 1010-4, 1010-5, 1010-6, 1010-7, 1010-8, 1015-2, 1015-3, 1015-4, 1015-5, 1015-6, 1015-7, 1015-14

---

# Index

- Communication with A/D converter, 1010-8, 1015-7
- Specifications, 510-5
- Aux 1
  - Temp, 820-2
  - Temp sensor, 850-12
  - Temp too hot, 850-12
- Aux 2
  - Temp, 820-2
  - Temp sensor, 850-13
  - Temp too hot, 850-13
- Aux 3
  - Faulty fact cal, 850-15
  - Method changed, 830-5
  - Pressure, 820-3
- Aux 4
  - Faulty fact cal, 850-15
  - Method changed, 830-5
  - Pressure, 820-3
- Aux 5
  - Faulty fact cal, 850-15
  - Pressure, 820-3
- Aux 5 method changed, 830-5
- Aux module rev, 850-16
- Aux wrong module, 850-17
- Auxiliary channel frit, changing, 360-14
- Auxiliary EPC
  - IPB, 1320-30
  - Replacement, 360-1
- Auxiliary zone
  - Wiring harness, 1240-4
- Auxiliary zone wiring harness, 1240-4

## B

### B det

- Adjust failure, 850-21
- Bad cksum, 850-20
- Bad fact cal, 850-20
- Calib deleted, 830-6
- Config changed, 830-3
- Electrometer, 850-4
- Invalid pid, 850-19
- Invalid type, 850-17
- I/O failure, 850-21
- Module rev, 850-15

- Temp sensor, 850-10
- Temp too hot, 850-10
- Type mismatch, 850-18
- Wrong module, 850-16

### B inlet

- Bad cksum, 850-20
- Bad fact cal, 850-20
- Calib deleted, 830-5
- Config changed, 830-4
- Flow cal fail, 830-8
- Invalid pid, 850-19
- Invalid type, 850-18
- I/O failure, 850-21
- Module rev, 850-16
- Temp sensor, 850-12
- Temp too hot, 850-11
- Wrong module, 850-17

### B TCD filament open, 850-5

### B TCD filament short, 850-6

### Back det

- Adjusting, 820-5
- Equib time, 820-5
- Flame out, 850-5
- Igniting, 820-5
- Shutdown, 820-6
- Temp, 820-2
- Waiting, 820-4

### Back inlet

- Purging, 820-8
- Temp, 820-2

### Bad mainboard errors, 860-1

### Bakeout

- ECD, 340-23, 341-25
- TCD, 330-23

### Battery replacement, 430-13

### BCD cables

- 03396-60560, 1010-6, 1010-7
- 03396-60570, 1010-12, 1015-10
- Connector, 1210-10, 1215-10
- G1530-60630, 1010-4, 1020-6

### Bead

- NPD, Replacement, 320-13

### Bead voltage, NPD, 320-4

### Bleed restrictor column, replacing, 260-4

### Boot ROM Checksum, 860-1

Bus Error, 860-1

## C

### Cables

03396-60530, 1010-10  
03396-60530 RS-232, 1010-11, 1015-9  
03396-60560,BCD, 1010-6  
03396-60560,BCD cables, 1010-7  
03396-60570,BCD, 1010-12, 1015-10  
03396-61010, 1010-11, 1015-9  
03396-61010,APG remote, 1010-7, 1015-6  
03396-61020, 1010-11, 1015-9  
03396-61020,APG remote, 1010-6, 1015-5  
24540-80012,modem, 1010-5  
24542M, 1010-2  
24542U,RS-232, 1010-9, 1015-8  
35900-60670, APG remote, 1010-4, 1010-12,  
1015-4, 1015-10  
8120-3446,GPIB, 1010-2, 1010-3, 1010-9  
82167-60003,INET, 1010-5, 1010-10  
Analog, 1020-1, 1210-2  
Automatic Liquid Sampler, 1020-3  
Auxiliary zone/Valve box wiring harness,  
1240-4  
BCD, 1020-6, 1210-10, 1215-10  
External connectors, 1210-1, 1215-1  
External Event, 1020-7, 1210-9, 1215-9  
G1290-60570, 1010-9, 1010-10, 1010-11, 1010-12,  
1015-8, 1015-9, 1015-10  
G1530-60560, 1020-2  
G1530-60560,Analog, 1010-12, 1015-10  
G1530-60570, 1010-7, 1015-6, 1020-1  
G1530-60570,Analog, 1010-6, 1010-8, 1010-11,  
1015-5, 1015-7, 1015-9  
G1530-60590, External events, 1010-4,  
1010-13, 1015-4  
G1530-60600,RS-232, 1010-2, 1010-3, 1010-4,  
1010-5, 1010-6, 1010-7, 1010-8  
G1530-60630,BCD, 1010-4  
G1530-60930,APG remote, 1010-2, 1010-3,  
1010-5, 1010-6, 1010-7, 1010-8, 1015-3, 1015-14  
G1530-61120,modem, 1010-5  
G2613-60590, 1020-4  
G2614-60610, 1020-5  
Inlet/detector harness, 1240-2

Modem, 1020-9

Modem/RS-232C, 1020-9

Remote, 1020-13, 1210-4, 1215-7

RS-232, 1020-3, 1210-7

Cables GPIB, 1210-11

Capillary Injection Port, 1310-11

Capillary Injection Port Sub-Assembly, 1310-11

Carbon dioxide for cryogenic cooling, 520-10

Carrier gas

Capillary columns, 520-2

NPD, 320-5

Catalyst Assembly IPB, 1350-12

Cautions, 110-1

Cell, TCD, replacement, 330-10

Channels, auxiliary, changing frits, 360-14

ChemStation

Communication with 6890 GC, 1010-2,  
1010-3, 1010-9, 1015-2, 1015-3, 1015-8, 1015-14

Communication with 7694 Headspace  
Sampler, 1010-9, 1015-8

Communication with Mass Selective  
Detector, 1010-3, 1015-3, 1015-14

Communication with modem, 1010-2, 1015-2  
Specifications, 510-5

Chromatographic troubleshooting, 1150-1

Circuitry

Heater driver, 1220-9

Inlet fan, oven flap, oven cryo, 1220-7

Valve driver, 1220-8

Cleaning

ECD, 340-23, 341-25

FID, 310-23

NPD, 320-30

TCD, 330-23

Col 1 config changed, 830-4

Col 2 config changed, 830-4

Collector

FID, cleaning, 310-23

FID, Replacement, 310-9

NPD, cleaning, 320-30

NPD, Replacement, 320-16

Column Liners, IPB, 1310-5

Column, installing, 240-28

Comm

Abnormal break, 830-7

---

# Index

- Data error, 830-6
- Data overrun, 830-6
- Computer, 510-5
- Computer specifications, 510-5
- Computer, non-Agilent
  - Specifications, 510-5
- Configuration
  - Power, 1230-1
  - Valves, 1140-1
- Connectors
  - AC power board, 1230-5
  - Analog, 1210-2
  - BCD, 1210-10, 1215-10
  - External, 1210-1, 1215-1
  - GPIB, 1210-11
  - INET, 1210-12
  - Main board, 1220-2, 1225-3
  - Remote, 1210-4, 1215-7
  - RS-232, 1210-7
  - Sampler injector or tray, 1210-14
- Consumables
  - Column liners, 1310-5
  - Packed column supplies, 1310-18
- Cool On-Column inlet
  - EPC manifold IPB, 1310-21
  - EPC manifold replacement, 230-8
  - Heater/Sensor replacement, 230-7
  - IPB, 1310-19
  - Leak points, 230-19
  - Leak test, 230-10
  - Replacement, 230-4
  - Theory of Operation, 230-1
- Cooling requirements, GC, 510-1
- Copper tubing recommendations, 520-6
- Covers
  - Automatic Liquid Sampler tray bracket, removal, 410-11
  - Injection port fan, removal, 410-16
  - Inlet carrier cover, removal, 410-12
  - IPB, 1330-1
  - Removing, 410-3
- Cryo assembly
  - Replacing the CO<sub>2</sub> or LN<sub>2</sub> inlet, 240-18
- Cryo circuitry diagram, 1220-7
- Cryo valve

- Installation, 420-5
- Replacement, 420-7
- Cryogenic cooling
  - Carbon dioxide, 520-10
  - Nitrogen, 520-12
  - Requirements, 520-9

## D

- DC voltages, Test points, 1220-1
- Detector
  - ECD, 340-1, 341-1
  - FID, 310-1
  - Gas recommendations, 520-1, 520-2
  - NPD, 320-1
  - TCD, 330-1
  - Wiring harness, 1240-2
- Detector interface board
  - ECD, Replacement, 340-16, 341-20
  - FID, Replacement, 310-20
  - NPD, Replacement, 320-28
  - TCD, Replacement, 330-21
- Diagnostics, 240-38
  - ECD, 340-17, 341-22
  - FID, 310-22
  - Shutdown behavior, 240-38
- Diagnostics mode, 820-8
- Divide by Zero, 860-1
- DMA FPGA Failure, 860-1
- DRAM Failure, 860-1
- Drivers, valves
  - Circuitry diagram, 1220-8
  - Installation, 1130-5
  - IPB, 1350-4
  - Types, 1130-4
- Driver, heater, ircuitry diagram, 1220-9

## E

- ECD
  - Cleaning, 340-23, 341-25
  - Diagnostics, 340-17, 341-22
  - Flow manifold replacement, 340-11
  - Frequency test, 340-17, 341-22
  - Interface board replacement, 340-16, 341-20
  - Leak test, 340-18, 341-23

---

# Index

- Replacing the entire detector, 340-4, 341-3
  - Signal board removal, 340-14, 341-18
  - Theory of operation, 340-1, 341-1
  - Wipe test, 340-24, 341-26
  - EEPROM
    - Replacement, 430-11
    - Troubleshooting, 860-1
  - Electrical
    - AC power board IPB, 1360-4
    - AC power supply IPB, 1360-2
    - Main circuit board IPB, 1360-6, 1360-8
    - Pneumatics board IPB, 1360-11
    - Shock hazard, 110-2
  - Electrical requirements
    - 6890 GC, 530-1
    - Cord, 530-2
    - Maximum consumption, 530-2
    - Oven type, 530-2
    - Voltages, 530-2
  - Electrometer
    - ECD, Replacement, 340-14, 341-18
    - FID, Replacement, 310-13
    - NPD, Replacement, 320-19
  - Electron Capture Detector
    - Gas recommendations, 520-1, 520-2
  - Electron Capture detector See ECD
  - Electron Capture Detector (ECD) Manifold
    - Assembly, 1320-4
  - Electronic Pressure Control See EPC
  - Electronics
    - AC power board circuitry, 1230-7
    - AC power board connector circuitry, 1220-6
    - Auxiliary zone/Valve box, 1240-4
    - Cover, 410-3
    - External connectors, 1210-1, 1215-1
    - Heater driver circuitry, 1220-9
    - Inlet fan, oven flap, oven cryo circuitry, 1220-7
    - Inlet/Detector wiring harness, 1240-2
    - Main board connectors, 1220-2, 1225-3
    - Valve driver circuitry, 1220-8
  - Environmental requirements, GC
    - Humidity, 510-1
    - Temperature, 510-1
  - EPC
    - Cool On-Column inlet, 230-2
    - ECD, 340-2, 341-2
    - ECD flow manifold replacement, 340-11
    - FID, 310-1
    - FID flow manifold replacement, 310-14
    - FPD flow manifold replacement, 350-24
    - Gas saver mode, 210-4
    - NPD, 320-1
    - NPD flow manifold replacement, 320-20
    - Pneumatics board replacement, 430-14
    - Purged/Packed inlet, 220-1
    - Split mode theory, 210-2
    - Splitless mode theory, 210-3
    - TCD, 330-5
    - TCD flow manifold replacement, 330-13
  - EPC inlets, performing the leak test, 210-21
  - EPC inlet, gas recommendations, 520-1, 520-2, 520-4
  - EPC manifold
    - Cool On-Column inlet, 1310-21
    - Cool On-Column inlet, replacing, 230-8
    - Gas supply fitting, Split/Splitless inlet, 210-18
    - Leak test, 210-26, 220-15, 230-15
    - Programmed Temperature Vaporization Inlet IPB, 1310-24
    - Purged/Packed inlet IPB, 1310-14, 1310-34
    - Purged/Packed inlet, replacing, 220-8
    - Replacing, 250-14
    - Replacing the Filter, 250-19
    - Split/Splitless inlet IPB, 1310-5
    - Split/Splitless inlet, replacing, 210-17
    - Volatiles Interface IPB, 1310-30
  - EPC module, correcting leaks, 210-29
  - Equilibration time, NPD, 320-4
  - Errors, bad mainboard, 860-1
  - Exception Vector, 860-1
  - External connectors, 1210-1, 1215-1
  - External device, 820-8
  - External Events
    - Cable, 1020-7
    - Connector, 1210-9, 1215-9
- F**
- F det

- Adjust failure, 850-21
- Bad cksum, 850-20
- Bad fact cal, 850-20
- Calib deleted, 830-6
- Config changed, 830-3
- Electrometer, 850-4
- Invalid pid, 850-19
- Invalid type, 850-17
- I/O failure, 850-21
- Makeup gas, 820-3
- Module rev, 850-15
- Temp sensor, 850-10
- Temp too hot, 850-9
- Type mismatch, 850-18
- Wrong module, 850-16
- F inlet
  - Bad cksum, 850-20
  - Bad fact cal, 850-20
  - Calib deleted, 830-5
  - Config changed, 830-4
  - Flow cal fail, 830-8
  - Invalid pid, 850-19
  - Invalid type, 850-18
  - I/O failure, 850-21
  - Module rev, 850-16
  - Temp sensor, 850-11
  - Temp too hot, 850-10
  - Wrong module, 850-17
- F TCD filament open, 850-5
- F TCD filament short, 850-6
- Fan
  - Injection port cover, removal, 410-16
  - Injection port, removal, 410-16
  - Motor replacement, 420-15
  - Pneumatics area, removal, 410-18
  - Replacement, 420-14
- Fast heating oven, 510-5
- Fatal errors, 860-1
- Fault 200—Pneumatics shutdown: faulty board, 850-1
- Fault 201—Pneumatics shutdown: faulty board, 850-1
- Fault 203—Signal DSP faulty, 850-2
- Fault 204—Sig DSP ROM broken, 850-2
- Fault 205—Sig DSP RAM broken, 850-3
- Fault 206—Sig DSP registers, 850-3
- Fault 207—Sig DSP data corrupt, 850-3
- Fault 208—Signal path test failed, 850-4
- Fault 209—Signal path test failed, 850-4
- Fault 210—Signal path test failed, 850-4
- Fault 211—Signal path test failed, 850-4
- Fault 212—Front electrometer out of specification, 850-4
- Fault 213—Back electrometer out of specification, 850-4
- Fault 214—Front detector flame out, 850-5
- Fault 215—Back detector flame out, 850-5
- Fault 216—Front TCD filament open, 850-5
- Fault 217—Back TCD filament open, 850-5
- Fault 218—Front TCD filament shorted, 850-6
- Fault 219—Back TCD filament shorted, 850-6
- Fault 220—Heater over current, 850-6, 850-8
- Fault 222—Oven thermal shutdown, 850-8
- Fault 223—Oven thermal shutdown, 850-9
- Fault 224—Oven thermal shutdown, 850-9
- Fault 225—Front detector thermal shutdown, 850-9
- Fault 226—Front detector thermal shutdown, 850-10
- Fault 227—Back detector thermal shutdown, 850-10
- Fault 228—Back detector thermal shutdown, 850-10
- Fault 229—Front inlet thermal shutdown, 850-10
- Fault 230—Front inlet thermal shutdown, 850-11
- Fault 231—Back inlet thermal shutdown, 850-11
- Fault 232—Back inlet thermal shutdown, 850-12
- Fault 233—Aux 1 thermal shutdown, 850-12
- Fault 234—Aux 1 thermal shutdown, 850-12
- Fault 235—Aux 2 thermal shutdown, 850-13
- Fault 236—Aux 2 thermal shutdown, 850-13
- Fault 237—No line interrupt thermal shutdown, 850-13
- Fault 238—Faulty line interrupt thermal shutdown, 850-14
- Fault 239—Mux ADC thermal shutdown, 850-14
- Fault 240—Mux ADC offset value, 850-14
- Fault 241—Line sense reading thermal shutdown, 850-14

- Fault 242—Pneu aux module invalid constants, 850-15
- Fault 243—Pneu aux module invalid constants, 850-15
- Fault 244—Pneu aux module invalid constants, 850-15
- Fault 245—Front det module: obsolete EEPROM, 850-15
- Fault 246—Back det module: obsolete EEPROM, 850-15
- Fault 247—Front inlet module: obsolete EEPROM, 850-16
- Fault 248—Back inlet module: obsolete EEPROM, 850-16
- Fault 249—Pres aux module: obsolete EEPROM, 850-16
- Fault 250—Front det: non-det module, 850-16
- Fault 251—Back det: non-det module, 850-16
- Fault 252—Front inlet: non-inlet module, 850-17
- Fault 253—Back inlet: non-inlet module, 850-17
- Fault 254—Non-aux module in pneu aux position, 850-17
- Fault 255—Front detector: invalid det module, 850-17
- Fault 256—Back detector: invalid det module, 850-17
- Fault 257—Front inlet: invalid inlet module, 850-18
- Fault 258—Back inlet: invalid inlet module, 850-18
- Fault 259—Front detector: det board not the same as module, 850-18
- Fault 260—Back detector: det board not the same as module, 850-18
- Fault 261—Host communication: MIO board defective, 850-18
- Fault 262—Host communications: RS232 defective, 850-18
- Fault 263—Host communications: GPIB defective, 850-19
- Fault 264—Sampler communications: RS232 defective, 850-19
- Fault 265—Front inlet: invalid pids, 850-19
- Fault 267—Front detector: invalid pids, 850-19
- Fault 268—Back detector: invalid pids, 850-19
- Fault 269—Pneu aux module: invalid pids, 850-19
- Fault 270—Front inlet: invalid module checksum, 850-20
- Fault 271—Back inlet: invalid module checksum, 850-20
- Fault 272—Front detector: invalid module checksum, 850-20
- Fault 273—Back detector: invalid module checksum, 850-20
- Fault 274—Pneu aux module: invalid module checksum, 850-20
- Fault 275—Front inlet: invalid constants from factory, 850-20
- Fault 277—Front detector: invalid constants from factory, 850-20
- Fault 278—Back detector: invalid constants from factory, 850-20
- Fault 279—Pneumatics aux: invalid constants from factory, 850-21
- Fault 280—F inlet read/write failure, 850-21
- Fault 281—B inlet read/write failure, 850-21
- Fault 282—F det read/write failure, 850-21
- Fault 283—B det read/write failure, 850-21
- Fault 284—Pneu aux read/write failure, 850-21
- Fault 285—Front detector offset adjustment failed, 850-21
- Fault 286—Back detector offset adjustment failed, 850-21
- FID
  - Cleaning, 310-23
  - Collector, cleaning, 310-23
  - Diagnostics, 310-22
  - Flow manifold replacement, 310-14
  - Ignition, 310-22
  - Ignition problems, 310-22
  - Interface board replacement, 310-20
  - Jet, cleaning, 310-23
  - Replacing the collector, 310-9
  - Replacing the detector weldment, 310-8
  - Replacing the electrometer, 310-13
  - Replacing the entire detector, 310-4
  - Replacing the jet, 310-11
  - Theory of operation, 310-1
- Filter, replacing, 240-26
- Flame Ionization Detector

---

# Index

- Gas recommendations, 520-1
- Flame Ionization detector See FID
- Flame Ionization Detector (FID), 1320-6
  - Manifold Assembly, 1320-8
- Flame Photometric Detector, 1320-22
  - Covers and Electronics, 1320-28
  - Gas recommendations, 520-1, 520-2
  - Jet and Transfer Tube Assemblies, 1320-24
  - PMT and Bracket Assemblies, 1320-26
- Flapper
  - IPB, 1340-4
  - Motor replacement, 420-17
- Flow
  - Cool On-Column inlet (EPC), 230-2
  - Cool On-Column inlet (manual), 230-3
  - ECD, 340-2, 341-2
  - ECD (manual), 340-2
  - FID, 310-1
  - FID (manual), 310-2
  - NPD, 320-1
  - NPD (manual), 320-3
  - Purged/Packed inlet (EPC), 220-1
  - Purged/Packed inlet (manual), 220-3
  - Split mode (EPC), 210-2
  - Splitless mode (EPC), 210-3
  - Splitless mode (manual), 210-5, 210-6
  - TCD, 330-5
  - TCD (manual), 330-5
- Flow manifold
  - Locating leaks, 210-27, 240-43, 250-27
  - Replacement, ECD, 340-11
  - Replacement, FID, 310-14
  - Replacement, FPD, 350-24
  - Replacement, NPD, 320-20
  - Replacement, TCD, 330-13
- FPD flow manifold replacement, 350-24
- Frequency test, ECD, 340-17, 341-22
- Front det
  - Adjusting, 820-5
  - Air flow, 820-3
  - Anode gas, 820-3
  - Equib time, 820-5
  - Flame out, 850-5
  - H2 flow, 820-3
  - Igniting, 820-5

- Ref flow, 820-3
- Shutdown, 820-6
- Temp, 820-2
- Waiting, 820-4

Front inlet

- Flow, 820-3
- Pressure, 820-3
- Purging, 820-8
- Temp, 820-2

Fuses

- AC Power board, 1230-4
- Main board, 1220-4
- Oven heater, 1230-2, 1230-4

## G

- G1900A Purge and Trap, 1010-14, 1015-13
- Gang fitting restrictor or O-ring
  - Replacing, 240-25
- Gas hazards, 110-3
- Gas plumbing
  - Overview, 520-5
  - Pipe-thread connections, 520-7
  - Recommendations, 520-5
  - Two-stage pressure regulator, 520-7
- Gas recommendations
  - Capillary columns, 520-2
  - Detectors, 520-1, 520-2
  - Packed columns, 520-1
  - Purity, 520-4
- Gas sampling valve 0, 820-7
- Gas saver, 250-7, 820-3
- Gas saver mode, 210-4
- Gas supply fitting
  - Split/Splitless inlet, 210-18
- Gas supply tubing
  - Diameter, 520-6
  - Recommendations, 520-6
- Gas supply, purity recommendations, 520-4
- GC
  - Benchtop requirements, 510-2
  - Cooling requirements, 510-1
  - Dimensions, 510-5
  - Electrical requirements, 530-2
  - Electrical specifications, 510-5
  - Grounding, 530-1

- Heat specifications, 510-5
- Humidity range, 510-1
- Line voltage, 530-1
- Space requirement, 510-2
- Temperature range, 510-1
- Venting exhaust, 510-1
- Venting noxious gases, 510-2
- Weight, 510-5
- GC Automatic Liquid Sampler, 510-5
- GC system
  - Component specifications, 510-5
- General purpose valves, 1110-4
- GPIB connector, 1210-11
- GPIB defective, 850-19
- Grounding
  - 6890 GC, 530-1
  - GC, 530-1
- H**
- Headspace Sampler, 7694
  - Communication with 3395 and 3396B
    - Integrators, 1010-11, 1015-9
  - Communication with 3396B/C INET
    - Integrator, 1010-10
  - Communication with 6890 GC, 1010-9,  
1010-10, 1010-11, 1010-12, 1015-8, 1015-9, 1015-10
  - Communication with ChemStation, 1010-9,  
1015-8
  - Specifications, 510-5
- Headspace sampler, 7694
  - Communication with 3396B, 1010-11, 1015-9
- Heated zones, 420-1
- Heaters
  - Driver circuitry, 1220-9
  - Oven fuses, 1230-2
  - Oven shroud, 1230-3
  - Overcurrent, 850-6
  - Testing the oven coil, 420-4, 1230-9
  - Zone wattages, 420-1
- Heater/Sensor
  - Cool On-Column inlet, 230-7
  - Purged/Packed inlet, 220-5
  - Split/Splitless inlet, 210-11
  - TCD, 330-11
- Heater/sensor

- Replacing assembly, 250-11
- Host system, 820-8
- Humidity, operating ranges, 510-1
- Hydrocarbon trap description, 520-8
- Hydrogen
  - NPD
    - turning off during solvent peak, 320-4
- Hydrogen shutdown, 850-1

## I

- Illegal Instruction, 860-1
- Incorrect ROM #2, 860-1
- Incorrect ROM #3, 860-1
- INET
  - Cable, 82167-60003, 1010-5
  - Connectors, 1210-12
  - IPB, 1360-9, 1360-15
- Inlet
  - Correcting leaks, 240-45
  - Programmable Cool On-Column, 230-1
  - Purged/Packed, 220-1
  - Reinstalling, 210-11
  - Split/Splitless, 210-1
  - Time before shutdown, 840-3
  - Wiring harness, 1240-2
- Inlet Adapter, replacing, 240-27
- Inlet carrier cover, removal, 410-12
- Inlet carrier, removal, 410-13
- Inlet fan, circuitry diagram, 1220-7
- Inlet, EPC, gas recommendations, 520-1, 520-2,  
520-4
- Inlet, nonEPC, gas recommendations, 520-1,  
520-2, 520-4
- Inlet/Detector wiring harness, 1240-2
- Installing MIO card, 430-7
- Integrator
  - 3395A/3396B communication with 6890  
GC, 1010-6, 1010-11, 1015-5, 1015-9
  - 3395A/3396B communication with 7694  
Headspace, 1010-11, 1015-9
  - 3395A/3396B communication with ALS,  
1010-6, 1015-5
  - 3395B/3396C communication with 6890  
GC, 1010-7, 1015-6

---

# Index

3395B/3396C communication with ALS,  
1010-7, 1015-6  
3396B/C INET communication with 6890  
GC, 1010-5, 1010-10  
3396B/C INET communication with 7694  
Headspace, 1010-10  
Specifications, 510-5  
Invalid heater power, 830-1  
Invalid line sense, 850-14

## J

Jet  
FID, cleaning, 310-23  
FID, Replacement, 310-11  
NPD, cleaning, 320-31  
NPD, Replacement, 320-18  
Jets  
NPD, 320-5

## K

Keyboard removal, 410-17

## L

Leak test  
Cool On-Column inlet, 230-10  
Cool On-Column inlet leak points, 230-19  
ECD, 340-18, 341-23  
EPC manifold, 210-26, 220-15, 230-15  
Performing, 240-40  
Preparation, 210-19  
Programmed temperature vaporization  
inlet, 240-39  
Programmed temperature vaporization  
inlet leak points, 240-45  
PTV manifold, 240-42, 250-26, 270-7  
Purged/Packed inlet, 220-10  
Purged/Packed inlet leak points, 220-19  
Split/Splitless inlet, 210-19  
Split/Splitless inlet leak points, 210-30  
Line interrupt, 850-14  
Line voltage  
Power configuration, 1230-1  
Requirements, 530-1

## Liner

Replacing, 240-35  
Liners, Column, 1310-5  
Liquid carbon dioxide  
Requirements for cryogenic cooling, 520-10  
Liquid nitrogen  
Requirements for cryogenic cooling, 520-12  
Liquid sample valves, 1110-5

## M

Main Board  
Fuses, 1220-4  
IPB, 1360-6, 1360-8  
Replacement, 430-1  
Replacing EEPROM chips, 430-11  
Test points, 1220-1  
Main board connectors, 1220-2, 1225-3  
Main FPGA Failure, 860-1  
Mainboard  
Bad mainboard/Fatal errors, 860-1  
Makeup gas, NPD, 320-5  
Manifold and plumbing assemblies, removing,  
240-22  
Manual detector  
ECD theory, 340-2  
FID theory, 310-2  
NPD theory, 320-3  
TCD theory, 330-5  
Manual inlets  
Cool OnColumn theory, 230-3  
Flow manifold carrier, 1330-6  
Purged/Packed theory, 220-3  
Split mode theory, 210-5  
Splitless mode theory, 210-6  
Mass Selective Detector  
Communication with 6890 GC, 1010-3,  
1015-3, 1015-14  
Communication with ChemStation, 1010-3,  
1015-3, 1015-14  
Specifications, 510-5  
Metal covers, 1330-2  
Microcell Electron Capture Detector ( $\mu$ ECD),  
1320-18  
Manifold Assembly, 1320-20  
MIO

---

# Index

- Connectors, 1210-12
- INET connectors, 1210-12
- IPB, 1360-9, 1360-15
- MIO board defective, 850-18
- MIO card
  - Installation, 430-7
  - Removal, 430-4, 430-23
- Modem
  - Cable, 1020-9
  - Communication with 6890 GC, 1010-5
- Modem cables
  - 24540-80012, 1010-5
  - G1530-61120, 1010-5
- Modular Input/Output card See MIO card
- Module
  - Valve leaks, 240-43
- Moisture trap description, 520-8
- Molecular sieve trap description, 520-8
- Motor
  - Oven fan replacement, 420-15
  - Oven flap replacement, 420-17
- Multiposition valve, 820-7
- Mux ADC offset value, 850-14

## N

- Nitrogen for cryogenic cooling, 520-12
- Nitrogen Phosphorus Detector (NPD), 1320-10
- Nitrogen Phosphorus Detector See NPD
- Nitrogen-Phosphorus Detector, 520-1, 520-2
- Nitrogen-phosphorus detector
  - bead voltage, 320-4
  - carrier and makeup flows, 320-5
  - equilibration time, 320-4
  - hardware problems, 320-7
  - hydrogen off, 320-4
  - jets, 320-5
  - solvent peak
    - turning hydrogen off, 320-4
- Nitrogen-Phosphorus Detector (NPD)
  - Manifold Assembly, 1320-12
- No 512Hz Interrupt, 860-1
- No line interrupt, 850-13
- No mux ADC response, 850-14
- Non-Agilent PC
  - Communication with 6890 GC, 1010-4, 1015-4

- NonEPC inlet, gas recommendations, 520-1, 520-2, 520-4
- NPD
  - Active element replacement, 320-13
  - Bead replacement, 320-13
  - Cleaning, 320-30
  - Collector replacement, 320-16
  - Collector, cleaning, 320-30
  - Detector replacement, 320-10
  - Electrometer replacement, 320-19
  - Flow manifold replacement, 320-20
  - Interface board replacement, 320-28
  - Jet replacement, 320-18
  - Jet, cleaning, 320-31
  - Theory of operation, 320-1

## O

- Operating instructions
  - Split mode, 250-1
  - Splitless mode, 250-2
- Oven
  - Cryo circuitry, 1220-7
  - Fan motor replacement, 420-15
  - Fan replacement, 420-14
  - Fast rates, 420-1
  - Flap motor replacement, 420-17
  - Heater fuses, 1230-2, 1230-4
  - Heater shroud, 1230-3
  - IPB, 1340-2
  - Power requirements, 510-5
  - Ramp rates, 420-1
  - Sensor replacement, 420-13
  - Temperature troubleshooting, 420-3
- Oven exhaust deflector, 510-2
- Oven flapper
  - Circuitry diagram, 1220-7
  - IPB, 1340-4
- Oven heater, testing resistance, 1230-9
- Oven sensor missing, 830-1
- Oven shroud
  - Replacement, 420-10
  - Sensor replacement, 420-13
- Oven temp, 820-1
  - Sensor, 850-9
  - Too cool, 850-9

---

# Index

Too hot, 850-8  
Overview, GC system components, 510-5  
Oxygen trap description, 520-8

## P

P aux bad fact cal, 850-21  
P aux calib deleted, 830-6  
PCB pneumatics board replacement, 430-14  
PCM, replacing, 270-2  
PCOC See Cool On-Column inlet  
Personal computer specifications, 510-5  
Pipe-thread connections, 520-7  
Plastic covers, 1330-4  
Plumbing assemblies, replacing, 240-23  
Pneu aux  
    Bad cksum, 850-20  
    Invalid pid, 850-19  
    I/O failure, 850-21  
Pneu board FPGA, 850-1  
Pneumatics  
    Direct mode, 250-5  
    Fan removal, 410-18  
    Split mode, 250-1  
    Splitless mode, 250-2  
Pneumatics board, 850-1  
    Replacement, 430-14  
Pneumatics Control Module  
    Replacement procedures, 270-2  
    Theory of Operation, 270-1  
Power board  
    Circuitry diagram, 1230-7  
    Connector circuitry diagram, 1220-6  
Power cord  
    Requirements, 530-2  
    Terminations, 530-4  
Power on in progress, 820-9  
Power requirements  
    6890 GC, 530-1  
    Cord, 530-2  
    Maximum consumption, 530-2  
    Oven type, 530-2  
    Voltages, 530-2  
Power supply  
    AC board connectors, 1230-5  
    AC board IPB, 1360-4

Fuses, 1230-4  
IPB, 1360-2  
Oven heater fuses, 1230-2  
Oven heater shroud, 1230-3  
Voltage configuration, 1230-1  
Power transformer replacement, 430-17  
Pre run, 250-4  
Pressure check, valves, 1150-5  
Pressure programming and gas saver, 250-7  
Pressure regulators  
    Connecting to gas tubing, 520-7  
    Pipe-thread connections, 520-7  
    Recommendations, 520-7  
Printer specifications, 510-5  
Programmable Cool On-Column inlet See Cool On-Column inlet  
Programmed Temperature Vaporization Inlet  
    IPB, 1310-22  
    Manifold IPB, 1310-24  
    Theory of Operation, 240-1  
Programmed temperature vaporization inlet  
    Leak points, 240-45  
    Leak test, 240-39  
Proportional control valve, troubleshooting, 840-4  
PTV  
    Replacing with cooling assembly, 240-16  
PTV inlet, replacing, 240-13  
PTV LCO<sub>2</sub> CryoAssy, 1310-26  
PTV LN<sub>2</sub> CryoAssy, 1310-27  
PTV manifold, leak test, 240-42, 250-26, 270-7  
PTV thermocouple PCB, replacing, 240-25  
Pulsed split mode, 210-2  
Pulsed splitless mode, 210-4, 240-8  
Purged/Packed inlet  
    Consumables, 1310-18  
    EPC manifold replacement, 220-8  
    Heater/sensor, 220-5  
    IPB, 1310-12  
    Leak points, 220-19  
    Leak test, 220-10  
    Manifold IPB, 1310-14, 1310-34  
    Replacement, 220-4  
    Theory of operation, 220-1

## R

Radioactivity test, ECD, 340-24, 341-26  
Ramp rates, 420-1  
Rate, fast oven, 420-1  
Regular oven, power requirements, 510-5  
Remote cable, 1020-13, 1210-4, 1215-7  
    Signal descriptions, 1210-5, 1215-7  
Remote start-stop cable See APG remote start-stop cable  
Replacement Procedures, 210-8, 240-13  
Replacement procedures  
    Pneumatics Control Module, 270-2  
Resistances  
    Heater coil, 420-4, 1230-9  
    Temperature sensors, 1240-7  
ROM chips  
    Replacement, 430-11  
ROM #2 Checksum, 860-1  
ROM #2 wrong version, 860-1  
ROM #3 Checksum, 860-1  
ROM #3 wrong version, 860-1  
RS-232  
    03396-60530,cable, 1010-11, 1015-9  
    24542M,cable, 1010-2  
    24542U cable, 1010-9  
    Connector, 1210-7  
    G1530-60600 cable, 1010-2, 1010-3, 1010-4, 1010-5, 1010-6, 1010-7, 1010-8, 1020-3  
RS-232 cable, signal descriptions, 1210-8, 1215-6  
RS-232 defective, 850-18

## S

Sampler  
    Abnormal comm, 830-7  
    ALS interface board, 430-23  
    Data error, 830-7  
    Data overrun, 830-7  
Sampler RS-232 defect, 850-19  
Sampling end setpoint, 250-3  
    Direct mode, 250-5  
Sensor  
    Oven sensor replacement, 420-13  
    Purged/Packed inlet, 220-5  
    TCD, 330-11

Sensors, temperature, resistance, 1240-7  
Septum head  
    Installing, 240-3  
    Removing/replacing, 240-33  
Septumless head  
    Cleaning, 240-30  
    Removing, 240-29  
Septum, replacing, 240-35  
Shock hazard, 110-2  
Shroud  
    Heater coil resistance, 1230-9  
    Oven, 1230-3  
Shutdown 10—Back detector fuel gas shutdown, 840-6  
Shutdown 11—Back detector air/ref shutdown, 840-6  
Shutdown 12—Back detector makeup shutdown, 840-6  
Shutdown 13—Pres aux 3 shutdown, 840-6  
Shutdown 14—Pres aux 4 shutdown, 840-6  
Shutdown 15—Pres aux 5 shutdown, 840-6  
Shutdown 16—Multiposition valve is not switching, 840-6  
Shutdown 17—Can't reach setpoint of multipos valve, 840-8  
Shutdown 1—Oven shut off, 840-1  
Shutdown 2—Cryo shut off, 840-2  
Shutdown 3—Front inlet pressure shutdown, 840-3  
Shutdown 4—Front inlet flow shutdown, 840-4  
Shutdown 5—Back inlet pressure shutdown, 840-4  
Shutdown 6—Back inlet flow shutdown, 840-4  
Shutdown 7—Front detector fuel gas shutdown, 840-5  
Shutdown 8—Front detector air/ref shutdown, 840-5  
Shutdown 9—Front detector makeup shutdown, 840-5  
Shutdown, time required for inlet to, 840-3  
Sig 1 buffer full, 830-2  
Sig 2 buffer full, 830-2  
Sig DSP  
    Data corrupt, 850-3  
    RAM broken, 850-3

---

# Index

- Registers, 850-3
- ROM broken, 850-2
- Signal 1/Signal 2 connectors, 1210-2
- Signal board, ECD, 340-14, 341-18
- Signal data loss, 830-3
- Signal descriptions
  - Remote, 1210-5, 1215-7
  - RS-232, 1210-8, 1215-6
- Signal DSP faulty, 850-2
- Solenoid valve, 250-4
- Solvent peak, NPD
  - turning hydrogen off, 320-4
- Solvent Vapor Exit Accessory
  - Leak testing, 260-7
  - Replacement Procedures, 260-2
  - Replacing the valve/fitting assembly, 260-2
  - Theory of Operation, 260-1
- Solvent Vent Mode, 240-9
  - Purge and cleanup, 240-12
  - Sample and vent, 240-9
  - Sample transfer, 240-11
- Space requirements
  - GC, 510-2
  - GC system configurations, 510-2
- Specifications
  - 3395 and 3396 Integrators, 510-5
  - 35900 Analog-to-Digital Converter, 510-5
  - 5972AMass Selective Detector, 510-5
  - 7694 Headspace Sampler, 510-5
  - ChemStation, 510-5
  - Computer, 510-5
  - Fast-heating oven, 510-5
  - GC, 510-5
  - Personal Computer, 510-5
  - Regular oven, 510-5
- Split modes, 240-3
- Split vent trap filter cartridge, replacement, 210-15
- Split vent trap, replacing, 210-13
- Splitless mode
  - Pulsed, 240-8
  - Volatiles Interface, 250-2
- Splitless Modes, 240-6
- Split/Splitless inlet
  - EPC manifold replacement, 210-17

- Gas saver mode, 210-4
- Heater/Sensor replacement, 210-11
- IPB, 1310-2, 1310-7, 1310-8
- Leak points, 210-30
- Leak test, 210-19
- Manifold IPB, 1310-5
- Replacement, 210-8
- Theory of operation, 210-2
- Start/Stop remote cable, 1210-4, 1215-7
- Static RAM Failure, 860-1
- SVE assembly, leak testing, 260-9
- System components specifications, 510-5

## T

### TCD

- Cleaning, 330-23
- Flow manifold replacement, 330-13
- Heater/sensor, 330-11
- Interface board replacement, 330-21
- Replacing the cell, 330-10
- Replacing the entire detector, 330-7
- Theory of operation, 330-1
- Teflon ferrule, replacing, 240-32
- Teflon tape, 520-7
- Temperature
  - Liquid sample valve ranges, 1110-5
  - Operating ranges, 510-1
  - Oven troubleshooting, 420-3
  - Ramps, 420-1
  - Sensor resistances, 1240-7
- Test points, 1220-1
- Theory of operation
  - Cool On-Column inlet, 230-1
  - ECD, 340-1, 341-1
  - FID, 310-1
  - NPD, 320-1
  - Pneumatics Control Module, 270-1
  - Programmed Temperature Vaporization
    - Inlet, 240-1
  - Purged/Packed inlet, 220-1
  - Solvent Vapor Exit Accessory, 260-1
  - Split/Splitless inlet, 210-2
  - TCD, 330-1
  - Volatiles Interface, 250-1
- Thermal Conductivity Detector

- Gas recommendations, 520-1, 520-2
- Thermal Conductivity detector See TCD
- Thermal Conductivity Detector (TCD), 1320-14
  - Manifold Assembly, 1320-16
- Tools required, 120-1
- Top insert assembly, replacement, 210-9
- Transformer, replacement, 430-17
- Traps
  - Activated charcoal, 520-8
  - Hydrocarbon trap, 520-8
  - Moisture, 520-8
  - Molecular sieve, 520-8
  - Overview, 520-8
  - Oxygen, 520-8
  - Split Vent, 510-2
- Tray bracket, removal, 410-11
- Tri-column assembly
  - Leak testing, 260-7
  - Replacing, 260-5
- Troubleshooting
  - 24V pneu valve drive, 820-7
  - Valves, 1150-1
- Tubing for gas supplies, 520-6
- Two-stage pressure regulators, 520-7

## V

- Valco valve 18900F
  - IPB, 1350-10
  - Overview, 1110-2
- Valve actuator air requirements, 520-13
- Valve Actuator Assembly, 1350-8
- Valve box
  - Installation, 1120-1, 1120-8
  - IPB, 1350-2
  - Wiring harness, 1240-4
- Valve driver IPB, 1350-4
- Valve fitting assembly
  - Solvent Vapor Exit Accessory, 260-2
- Valves
  - Actuator alignment, 1130-10
  - Actuator installation, 1130-1
  - Actuator IPB, 1350-6
  - Bodies, 1110-3
  - Configuration diagrams, 1140-1
  - Driver circuitry, 1220-8

- Driver installation, 1130-5
- Drivers, 1130-4
- General purpose, 1110-4
- IPB, 1350-10, 1350-12
- Liquid sample, 1110-5
- Pressure check, 1150-5
- Proportional control
  - Troubleshooting, 840-4
- Rotors, 1110-3
- Temperature ranges, LSV, 1110-5
- Troubleshooting, 1150-1
- Valco 18900F, 1110-2
- Valve box installation, 1120-1, 1120-8
- Venting requirements
  - Detectors, 510-2
  - Oven, 510-1
  - Oven exhaust reflector, 510-1
  - Split/splitless inlet, 510-2
- Volatiles Interface
  - Calibrating, 250-18
  - Correcting leaks, 250-29
  - Direct mode, 250-4
  - IPB, 1310-26
  - Leak checking, 250-22
    - Materials needed, 250-22
  - Leak testing, 250-22
  - Manifold IPB, 1310-30
  - Potential leak areas, 250-30
  - Replacement procedures, 250-8
  - Replacing or cleaning, 250-8
  - Split mode, 250-1
  - Splitless mode, 250-2
  - Theory of Operation, 250-1
- Voltages
  - Power Configuration, 1230-1
  - Test points, 1220-1

---

# Index

## W

Warning 100—Oven sensor missing, 830-1  
Warning 101—Invalid heater power for front, 830-1  
Warning 102—Invalid heater power for back, 830-1  
Warning 103—Sig 1 buffer full, 830-2  
Warning 104—Sig 2 buffer full, 830-2  
Warning 105—Analog out data loss, 830-3  
Warning 106—Non-recoverable data loss, 830-3  
Warning 107—Front det config changed, 830-3  
Warning 108—Back det config changed, 830-3  
Warning 109—Front inlet config changed, 830-4  
Warning 110—Back inlet config changed, 830-4  
Warning 111—Column 1 config changed, 830-4  
Warning 112—Column 2 config changed, 830-4  
Warning 113—Aux 3 config changed, 830-5  
Warning 114—Aux 4 config changed, 830-5  
Warning 115—Aux 5 config changed, 830-5  
Warning 117—F inl calib deleted, 830-5  
Warning 118—B inlet calib deleted, 830-5  
Warning 119—F det calib deleted, 830-6  
Warning 120—B det calib deleted, 830-6  
Warning 121—P aux calib deleted, 830-6  
Warning 122—Host communications: data overrun, 830-6  
Warning 123—Host communications: data error, 830-6  
Warning 124—Host communications: abnormal break, 830-7  
Warning 125—Sampler communications: data overrun, 830-7  
Warning 126—Sampler communications: data error, 830-7  
Warning 127—Sampler communications: abnormal break, 830-7  
Warning 128—Front inlet sensor auto zero calib failed, 830-8  
Warning 129—Back inlet sensor auto zero calib failed, 830-8  
Warnings, 110-2  
Wattage  
  Heaters, 420-1  
Weight  
  GC, 510-2

GC system components, 510-2

Table of weights, 510-5

Weldment

  FID, Replacement, 310-8

Wipe test, ECD, 340-24, 341-26

## Z

Zero flow, 130-1

Zones, heated, 420-1