

# Index

---

## A

- a.c. power connection 15
- acquisition parameters 124
- Acquisition Setup dialog box 57
- adduct list
  - creating 109
  - opening 111
  - saving 110
- analog data 69
  - acquiring 75
- Analog Data dialog box 76
- Analogue Channel 10
- APCI 260
- APCI probe 3
  - heater
    - not working 219
    - replacing 206
  - installing 38
  - removing 17
  - stainless steel capillary, replacing 196
  - tip, cleaning 181
- APPI electrical connection 7
- assembling
  - source ion block 175
  - source T-Wave assembly 176
- atmospheric pressure chemical
  - ionization 260
- automated quantification 63
- Automatic Calibration Check dialog box 119
- automatic calibration check parameters 119
- Automatic Calibration dialog box 123
- Aux O/P connection 12

## B

- backing pump 3
- baseline, resetting 238
- biological hazard iii

## C

- calibration
  - checking 129
  - failure 127
  - peak matching, manually editing 131
  - performing 123
  - removing 118
  - report 129
  - saving 132
  - scan speed compensation 123
  - scanning 123
  - selecting the reference file 118
  - setting general parameters 120
  - setting parameters 119
  - starting 125
  - static 123
  - verifying 132
- Calibration Acquisition Setup dialog box 124
- Calibration Options window 126
- Calibration Parameters dialog box 120
- calibration reference files
  - editing 251
  - negative ion 254
  - positive ion 252
  - information 251
- calibration report 126, 129
- CE Int connection 12

- CE Profile dialog box 105
- centroid data 68, 80
- Charge State Mass dialog box 107
- chemical hazard iii
- chemical noise 223
- chromatogram, viewing in real-time 65
- cleaning
  - APCI probe tip 181
  - cone gas cone 172
  - corona discharge pin 180
  - ESI probe tip 178
  - extraction cone 174
  - gas exhaust port 172
  - ion block 174
  - isolation valve stem 175
  - sample cone 172
  - source T-Wave assembly 173
- collision energy
  - options 102
  - profile table 105
    - adding new entries 107
    - opening 108
    - saving 108
  - ramp
    - applying 86
    - controlling 238
    - setting-up 237
- Collision Gas In connection 14
- collision gas, controlling 235
- Collision Ramp dialog box 86, 237
- Com1 connection 12
- Comm For EPC connection 12
- communications status 70
- complete shutdown 228
- cone gas
  - connection 6
  - controlling 235
- cone gas cone, cleaning 172
- Cone Ramp dialog box 79, 236

- cone voltage ramp
  - applying 79
  - controlling 237
  - setting-up 236
- connections 10
  - front panel 4
  - rear panel 10
  - vacuum 15
- constant neutral loss mode 266
- contact closure connections 11
- contacting Waters 224
- continuum data 80
- controlling
  - collision energy ramp 238
  - cone gas 235
  - cone voltage ramp 237
  - desolvation gas 235
  - gas flows 235
  - nebulizer gas 235
  - syringe pump 235
- controls
  - embedded PC reset switch 4
  - front panel 4
  - Inject switch 9
  - Load switch 9
  - power switch 4
- corona discharge pin
  - cleaning 180
  - installing 37
  - removing 19
  - replacing 180

## D

- data
  - analog 69
  - centroid 68, 80
  - continuum 80
  - MCA 80

- profile 67
- SIR 68
- data acquisition
  - monitoring 65
  - pausing 70
  - starting 57
  - stopping 70
- daughter (product) ion mode 262
- Daughter Scan Function Editor 85
- desolvation gas
  - connection 6
  - controlling 235
- desolvation heater not working 219
- dialog boxes
  - Acquisition Setup 57
  - Analog Data 76
  - Automatic Calibration 123
  - Automatic Calibration Check 119
  - Calibration Acquisition Setup 124
  - Calibration Parameters 120
  - CE Profile 105
  - Charge State Mass 107
  - Collision Ramp 86, 237
  - Cone Ramp 79, 236
  - Display Calibration Graphs 129
  - Exclude Mass 101
  - Include Masses 97
  - Instrument Threshold Settings 66
  - Mass Measure Parameters 121
  - Modify Charge State 106
  - Quantify Samples 63
  - Readbacks 238
  - Scope Setup 236
  - Solvent Delay 74
  - Spectrum Real-Time Update 65
  - Start Sample List Run 60
  - Syringe Selection 235
- disassembling
  - source ion block 164

- source T-Wave assembly 162
- Display Calibration Graphs dialog box 129
- divert/injection valve 8

## **E**

- electrical connections
  - a.c. power 15
  - Analogue Channel 10
  - APPI 7
  - Aux O/P 12
  - CE Int 12
  - Com1 12
  - Comm For EPC 12
  - contact closure 11
  - Events I/P 11
  - Events O/P 11
  - Gas Fail 11
  - HV 7
  - mains power 15
  - MUX 7
  - Probe 7
  - Pump Relay 12
- electronic noise 223
- electrospray ionization 259
- embedded PC reset switch 4
- emergency shutdown 228
- ESCi multi-mode operation 3, 53
  - data acquisition 55
  - option key disk 53
  - preparing the instrument 53
  - tuning the instrument 54
- ESI 259
- ESI probe 3
  - installing 20
  - removing 35
  - stainless steel capillary, replacing 187
  - tip
    - cleaning 178

- replacing 178
- Events I/P connections 11
- Events O/P connections 11
- Exclude List 99
  - editing 101
- Exclude Mass dialog box 101
- Exhaust connection 14
- extraction cone, cleaning 174

## F

- filter elements
  - changing 145
- fitting
  - gas exhaust port 178
  - ion block 177
  - ion source enclosure 177
  - source T-Wave assembly 176
- flammable solvents operation hazard iv
- front panel
  - connections 4
  - controls 4
- Function Editors
  - Daughter Scan 85
  - Full Scan 77
  - MRM 88
  - Parent Scan 87
  - SIR 81
  - Survey Scan 88
- function list
  - adding a new function 73
  - changing the order of functions in 74
  - copying an existing function 74
  - modifying an existing function 73
  - opening a saved list 77
  - removing a function 74
  - saving 76
  - setting a solvent delay 74
  - setting the maximum retention time 74

- setting-up 71
- functions
  - MRM 88
  - SIR 81
  - survey 88
- fuses
  - main system printed circuit board 219
  - mains power 15

## G

- gas connections
  - Collision Gas In 14
  - cone gas 6
  - desolvation gas 6
  - Exhaust 14
  - Gas In 13
  - nebulizer gas 6
- gas exhaust port
  - cleaning 172
  - fitting 178
  - removing 156
- Gas Fail connection 11
- gas flow rates 16
- gas flows, controlling 235
- Gas In connection 13
- gas-ballasting, rotary pump 141

## H

- hazards
  - biological iii
  - chemical iii
  - flammable solvents operation iv
  - high temperature v
  - high voltage iv
  - solvent leakage iv
- high LC backpressure 216
- high noise levels 222

high temperature hazard v  
high voltage hazard iv  
HV electrical connection 7

## I

IEEE communication errors 222  
Include List 95  
    editing 97  
Include Masses dialog box 97  
Inject switch 9  
Inlet Configuration Wizard 33  
installing  
    APCI probe 38  
    corona discharge pin 37  
    ESI probe 20  
instrument data thresholds 66  
instrument name, selecting 247  
instrument parameter files 239  
    creating new 240  
    opening 240  
Instrument Threshold Settings dialog box 66  
insufficient vacuum 217  
intended use xiv  
inter scan delay, selecting 236  
ion block  
    cartridge heater, replacing 182  
    cleaning 174  
    fitting 177  
    removing 157  
ion counting threshold 68  
Ion Mode menu 233  
ion modes  
    daughter 262  
    parent 264  
    precursor 264  
    product 262  
ion optics 260  
ion source enclosure

    fitting 177  
    removing 157  
ion source heater, not working 219  
ionization mode, selecting 233  
isolation valve 3, 23, 41, 152  
    stem, cleaning 175  
items exposed to solvent 269

## K

key disk, ESCi option 53

## L

LEDs  
    Operate 7  
    Vacuum 7  
Load switch 9  
loss of communication with instrument 221  
low intensity peaks 215  
low LC backpressure 217

## M

mains power  
    connection 15  
    fuses 15  
maintenance  
    procedures 133  
    safety 134  
    schedule 133  
mass calibration 113  
    for electrospray 114  
    scan speed 113  
    scanning 113  
    static 113  
mass flow controllers 16  
mass measure parameters 121  
Mass Measure Parameters dialog box 121

## MassLynx

- ESCi option key disk 53
- software 4
- Tune window, opening 233
- MassLynx control system 3
- materials of construction 269
- maximum retention time 74
- MCA data 80
- mobile phases, common ingredients 270
- Modify Charge State dialog box 106
- monitoring acquisitions 111
- MRM function 88
- MRM Function Editor 88
- MS Method Editor 71
  - toolbar 72
- MS operating modes 261
- MS to MS/MS switching 90
- MS/MS operating modes 262
- MS/MS scanning functions 84
- MS/MS to MS switching 93
- MS2 scan 87
- Multi Channel Analysis data 80
- multiple reaction monitoring mode 265
- multiple-sample acquisition 60
- MUX electrical connection 7

## N

- nebulizer gas
  - connection 6
  - controlling 235
- neutral gain scan 87
- neutral loss scan 87
- no ion beam 213
- no peaks in the Tune window 213
- noise
  - chemical 222
  - electronic 222

## O

- odor filter element, changing 145
- oil mist filter element, changing 145
- Operate LED 7
- operating modes
  - MS 261
  - MS/MS 262
- overnight shutdown 228

## P

- parameters
  - acquisition 124
  - mass measure 121
  - setting calibration 119
  - setting general calibration 120
- parameters automatic calibration check 119
- parent (precursor) ion mode 264
- Parent Scan Function Editor 87
- Peak Display menu 244
- peak display, customizing
  - colors 245
  - display grid 246
  - intensity display 246
  - numbers 245
  - trace line appearance 246
- peaks
  - changing gain 244
  - low intensity 215
  - selecting
    - for tuning 241
    - operating mode 242
    - span 243
    - tune mass 242
  - unsteady 215
- performance specifications 257
- Pirani gauge 3
- power switch 4

printing tune information 240

probe

APCI 3

connections 150

ESI 3

removing 149

Probe electrical connection 7

profile data 67

removing spikes from 69

pump

backing 3

turbomolecular 3

Pump Relay connection 12

## Q

Quantify Samples dialog box 63

## R

Readbacks dialog box 238

real-time viewing, chromatogram 65

rear panel connections 10

reference file, selecting 118

reference sample solutions, preparing 255

removing

APCI probe 17

corona discharge pin 19

ESI probe 35

gas exhaust port 156

ion block 157

ion source enclosure 157

replacing

APCI probe heater 206

APCI probe stainless steel capillary 196

corona discharge pin 180

ESI probe stainless steel capillary 187

ESI probe tip 178

ion block cartridge heater 182

report, calibration 126

reset switch, embedded PC 4

rotary pump

fuse failure 219

gas-ballasting 141

oil

changing 144

checking 143

## S

safety Information iii

safety procedures, maintenance 134

safety symbols x

sample cone

cleaning 172

fitting to instrument 178

removing 151

sample inlet 3

Sample List 60

Scan Report window 65

scan speed compensation calibration 123

scan time, selecting 236

scanning calibration 123

scans

MS2 87

neutral gain 87

neutral loss 87

Scope Setup dialog box 236

shutdown

complete 228

emergency 228

overnight 228

shutting down the instrument 228

SIR data 68

SIR function 81

SIR Function Editor 81

adding a function 83

- solvent
  - flammable hazard iv
  - items exposed to 269
  - leakage hazard iv
- Solvent Delay dialog box 74
- source ion block
  - assembling 175
  - disassembling 164
- source T-Wave assembly
  - assembling 176
  - cleaning 173
  - disassembling 162
  - fitting 176
  - removing 160
- source, cleaning 148
- Spectrum Real-Time Update dialog box 65
- spectrum, viewing in real-time 65
- Start Sample List Run dialog box 60
- starting up the instrument 225
- static calibration 123
- status display LEDs 7
- survey function 88
- Survey Scan Function Editor 88
  - Adduct page 109
  - Collision Energy page 102
  - Exclude page 99
  - Include page 95
  - MS to MSMS page 90
  - MSMS Template page 89
  - Survey page 89
- switches
  - Inject 9
  - Load 9
- switching
  - MS to MS/MS 90
  - MS/MS to MS 93
- syringe pump
  - controlling 235
  - setting-up 231

- Syringe Selection dialog box 235
- syringe type, selecting 235
- System Manager window 70

## T

- thresholds
  - instrument data 66
  - ion counting 68
- troubleshooting 211
  - APCI probe heater not working 219
  - chemical noise 223
  - desolvation heater not working 219
  - electronic noise 223
  - failure to recognize a probe 220
  - hardware 213
  - high LC backpressure 216
  - high noise levels 222
  - IEEE communication errors 222
  - instrument will not turn on 213
  - insufficient vacuum 217
  - ion mode fault 220
  - ion source heater not working 219
  - isolation valve 214
  - leaking nitrogen 218
  - loss of communication with
    - instrument 221
  - low LC backpressure 217
  - no ion beam 213
  - no peaks in the Tune window 213
  - ripple 221
  - rotary pump fuse failure 219
  - rotary pump oil in exhaust tubing 218
  - safety 211
  - system 211
  - unsteady or low intensity peaks 215
- tune information, printing 240
- tune parameters
  - changing 239



saving 239  
Tune window  
  Diagnostics page 247  
  EasyTune Source page 241  
  ESCi source page 54  
  T-WAVE page 248  
turbomolecular pump 3  
T-Wave  
  optics, manually controlling 247  
  theory 267

## U

unsteady peaks 215

## V

vacuum  
  connections 15  
  system 3  
Vacuum LED 7

## W

Waste connection 15  
Waters, contacting 224

## Z

zero level, resetting 238  
ZSpray ionization source 3